

Safety Self-Inspection Tag

Building:	Room:
Dullully.	NOOIII.

Contact Person: Year:

Inspected by:

Month	W1	W2	W3	W4	М
January					
February					
March					
April					
May					
June					
July					
August					
September					
October					
November					
December					

DAILY

Immediate Hazards – identify any observed hazards and address as appropriate, e.g.

- poor housekeeping
- o inadequate attire and/or PPE
- o food/drink/cosmetics in lab
- improper labeling of materials or unattended experiments

- o propped doors
- obstruction of exits, aisles, pull stations or emergency equipment
- o unsafe equipment or work practices

WEEKLY

 Eyewash – activate eyewash station to flush line and ensure proper operation; ensure area around station is clear

MONTHLY (as applicable):

- **Fire Extinguisher** confirm extinguisher is in place & verify it is fully charged (green zone on pressure gauge)
- First Aid/Spill kit check kit contents and replace items if required
- Fume hoods check certification of alarms & notify EHS if certification has expired. Notify PR if malfunctioning. Ensure work area & airflow is not obstructed
- BSCs check certification and notify EHS if certification has expired. Notify BSC manufacturer if BSC is malfunctioning. Ensure work area & airflow is not obstructed
- Chemical Storage inspect storage areas for leakage, expired/old chemicals, and separation of incompatibles.
 Ensure flammables are stored in flammable storage cabinets.
- Emergency Lighting check for proper functioning of emergency lighting and replace as necessary

PERIODICALLY/ANNUALLY (as applicable):

- Chemical Inventory confirm SDSs are available for all chemicals in the lab and online inventory is up to date
- Warning Signs confirm signage required for biosafety, radioactive materials, lasers, mechanical hazards, noise, chemical or electrical hazards are in place and visible